

## Computer requirements for the Aspel products

Aspel S.A.

Computer compatible with IBM (acc. to PN-EN 60950)

Aspel guarantees proper software's operation only on the computers with MS Windows environment, comply with specification mentioned below.

Product	Operating system	Number of ports		Additional computer configuration	Monitor's resolutions	Soft version
		Rs-232	USB			
HolCARD 24W Alfa System A712 (Product's code: 07-102-0048)	Windows 7 32/64bit Windows 8.1 32/64bit Windows 10 32/64bit	--	4	SD card reader CD/DVD R/W	1280x800 1280x1024 1360x768 1440x900 1600x900 1920x1080	8.05.00
HolCARD 24W Alfa System A703 (Product's code: 07-102-0046)	Windows 7 32/64bit Windows 8.1 32/64bit Windows 10 32/64bit	--	4	SD card reader CD/DVD R/W	1280x800 1280x1024 1360x768 1440x900 1600x900 1920x1080	8.05.00
HolCARD 24W Alfa System A800 (Product's code: 07-102-0032)	Windows 7 32/64bit Windows 8.1 32/64bit Windows 10 32/64bit	--	4	CF card reader CD/DVD R/W	1280x800 1280x1024 1360x768 1440x900 1600x900 1920x1080	8.05.00
HolCARD 24W Alfa System A812 (Product's code: 07-102-0033)	Windows 7 32/64bit Windows 8.1 32/64bit Windows 10 32/64bit	--	4	CF card reader CD/DVD R/W	1280x800 1280x1024 1360x768 1440x900 1600x900 1920x1080	8.05.00
HolCARD 24W Alfa System A903 (Product's code: 07-102-0034)	Windows 7 32/64bit Windows 8.1 32/64bit Windows 10 32/64bit	--	4	SD card reader CD/DVD R/W	1280x800 1280x1024 1360x768 1440x900 1600x900 1920x1080	8.05.00
HolCARD CR-07 Alfa System (Product's code: 07-105-0004)	Windows 7 32/64bit Windows 8.1 32/64bit Windows 10 32/64bit	--	3	CD/DVD R/W	17" 1024x768 (*)	3.10.00

## Computer requirements for the Aspel products

Aspel S.A.

CardioTEST Alfa System B612 (Product's code: 07-103-0027)	Windows 7 32/64bit Windows 8.1 32/64bit Windows 10 32/64bit	1	3	CD/DVD R/W, internal or external speaker	17" 1024x768 (*)	3.50.00
CardioTEST Alfa System CRG200 (Product's code: 07-103-0028)	Windows 7 32/64bit Windows 8.1 32/64bit Windows 10 32/64bit	1	3	CD/DVD R/W, internal or external speaker	17" 1024x768 (*)	3.50.00
CardioTEST Alfa System A500 (Product's code: 07-103-0018)	Windows 7 32/64bit Windows 8.1 32/64bit Windows 10 32/64bit	1	3	CD/DVD R/W, internal or external speaker	17" 1024x768 (*)	3.50.00
CardioTEKA - software (Product's code: 07-101-0073)	Windows 7 32/64bit Windows 8.1 32/64bit Windows 10 32/64bit	0 or 1 if ECG AsCARD device is equipped with RS	2	CD/DVD R/W, Ethernet and sound cards	17" 1024x768 (*)	5.04.00
CardioTEKA Amber (Product's code: 07-101-0074)	Windows 7 32/64bit Windows 8.1 32/64bit Windows 10 32/64bit	0	3	CD/DVD R/W	17" 1024x768 (*)	5.04.00
CardioTEKA Coral (Product's code: 07-101-0402)	Windows 7 32/64bit Windows 8.1 32/64bit Windows 10 32/64bit	0	3	CD/DVD R/W	17" 1024x768 (*)	5.04.00
AsTER	Windows 7 32/64bit Windows 8.1 32/64bit Windows 10 32/64bit	1 (optional + 1 for each next device)	2 (optional +1 for each CR07)	CD/DVD R/W	17" 1024x768 (*)	3.50.00
CardioTEKA Viewer	Windows 7 32bit	0 or 1 if ECG AsCARD device is equipped with RS	2	---	17" 1024x768 (*)	3.24.00
AsPRINTER	Windows 7 32bit	1	2	---	17" 1024x768 (*)	3.25.00
AsCOMMTEL	Windows 7 32bit	--	--	Ethernet card	17" 1024x768 (*)	1.00.00

(\*) – minimum size and resolution of the monitor

The Aspel S.A. is not responsible for not working and improper working of the Aspel's software as well as for any consequences from not working and improper working resulting from noncompliance with the operation manual, using unlicensed software or any other unsure software installed in user's PC where Aspel's software is installed, and presence of other malware on the computer where software of ASPEL is installed.

The Microsoft has discontinued support for MS WindowsXP and MS Windows8, and hence are not possible updates for these operating systems. Therefore, ASPEL stopped to provide assistance for the installation and operation of Aspel's software installed on aforementioned systems due to fact that there is no guarantee that these operating systems will function properly and stable. In consequence ASPEL software installed on these systems may not work properly. You are asked about changing the operating system into recommended and indicated in the above table for particular product.